



Sandy Lane Industrial Estate, Worksop, Notts S80 1TN
 Tel: 01909 722960 info@selectscales.co.uk

CALIBRATION CERTIFICATE

CUSTOMER
 Select Scales

MACHINE DETAILS
 Make: T-Scale
 Model: KC
 Capacity: 15kg
 Division: 0.002kg
 Serial no. 1234567
 Location:

Certificate No. SEL1001
 Test Date: 19th July 2022
 Next Test Due: July 2023

The following certification tests have been carried out using calibrated weights traceable to the National Physical Laboratory Standards of Mass, uncertainties quoted provide a level of confidence of approximately 95%.
 The certificate provided, is in accordance with the UK Weighing Federation Code of Practice.

Linearity and Hysteresis				
Load	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0kg	0.000kg	Nil		
2kg	2.000kg	Nil		
4kg	4.000kg	Nil		
6kg	6.000kg	Nil		
8kg	8.000kg	Nil		
10kg	10.000kg	Nil		
12kg	12.000kg	Nil		
15kg	15.000kg	Nil		

Eccentric Loading			
	2	3	
	1	4	
As Found Test			
	Test Point	Reading	Deviation
Pos. 1	5kg	5.000kg	Nil
Pos. 2	5kg	5.000kg	Nil
Pos. 3	5kg	5.002kg	+1
Pos. 4	5kg	5.000kg	Nil

Definitive Test			
	Test Point	Reading	Deviation
Pos. 1			
Pos. 2			
Pos. 3			
Pos. 4			

Repeatability (x3) tested at:				
Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
8kg	8.000kg	Nil		
8kg	8.000kg	Nil		
8kg	8.000kg	Nil		

Test Weight Information	
Cert no. SH10586	Weights Set 1 11 & T10

Remarks (To include weighing range if less than the maximum weighing capacity)	Tolerances (MPE) Class III
	0-500divs = ± 1 div 500-2000divs = ± 2 divs 2000-3000divs = ± 3 divs

Engineers Signature

Engineers Name (Print).....