



Sandy Lane Industrial Estate, Worksop, Notts S80 1TN  
Tel: 01909 722960 info@selectscales.co.uk

## CALIBRATION CERTIFICATE

**CUSTOMER**

SAMPLE

**MACHINE DETAILS**

Make: Dini Argeo  
Model: DFW  
Capacity: 6kg  
Division: 0.001kg  
Serial no. 11112222222  
Location: XYZ

**Certificate No. ABCD1234**

Test Date: 1st February 2025  
Next Test Due: February 2026

The following certification tests have been carried out using calibrated weights traceable to the National Physical Laboratory Standards of Mass, uncertainties quoted provide a level of confidence of approximately 95%.  
The certificate provided, is in accordance with the UK Weighing Federation Code of Practice.

Linearity and Hysteresis				
Load	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0kg	0.000kg	Nil		
1kg	1.000kg	Nil		
2kg	2.000kg	Nil		
3kg	3.000kg	Nil		
4kg	4.000kg	Nil		
5kg	5.000kg	Nil		
6kg	6.000kg	Nil		

Eccentric Loading			
	2	3	
	1	4	
As Found Test			
	Test Point	Reading	Deviation
Pos. 1	2kg	2.000kg	Nil
Pos. 2	2kg	2.000kg	Nil
Pos. 3	2kg	2.000kg	Nil
Pos. 4	2kg	2.000kg	Nil

Definitive Test			
	Test Point	Reading	Deviation
Pos. 1			
Pos. 2			
Pos. 3			
Pos. 4			

Repeatability (x3) tested at:				
Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
3kg	3.000kg	Nil		
3kg	3.000kg	Nil		
3kg	3.000kg	Nil		

Test Weight Information	
Cert no.	Weights
SAMPLE	

Remarks (To include weighing range if less than the maximum weighing capacity)	Tolerances (MPE)	Class III
	0-500divs = ± 1 div	
	500-2000divs = ± 2 divs	
	2000-3000divs = ± 3 divs	

Engineers Signature ..... Engineers Name (Print)..... (F).29/2