



1114

Select Scales

Weighing Equipment Sales and Services

Sandy Lane Industrial Estate, Worksop, Notts S80 1TN

Tel: 01909 722960 info@selectscales.co.uk



WEIGHING PERFORMANCE CERTIFICATE SAMPLE

CUSTOMER	Test Date: 15 th November 2026 Next Test Due: November 2027
-----------------	---

EQUIPMENT DETAILS				
MAKE	MODEL	SERIAL NO.	CAPACITY	TOLERANCE
Dini Argeo	DFW	86995	6500kg x 1kg	+/- 3 Divisions

<u>LINEARITY READINGS</u>		
Load Applied	As Found	Final Indications
0kg	0kg	
1000kg	1000kg	
2000kg	2000kg	
3000kg	3000kg	
4000kg	4000kg	
5000kg	5000kg	
6000kg	6000kg	

<u>REPEATABILITY READINGS</u>		
Load Applied	As Found	Final Indications
3000kg	3000kg	
3000kg	3000kg	
3000kg	3000kg	

Certificate no.PSLN004672 TEST EQUIPMENT USED TW035

Remarks

Engineers Signature Engineers Name (Print)... ..

This certificate is provided for internal purposes only